



California Medical Device Recall Information



Recall Name

CareFusion Corporation Recalls AirLife Infant Breathing Circuit Due to a Potential Leak in the System

Recall Date	Product Description	Recalling Firm	Recall Reason
5/29/12	AirLife Infant Breathing Circuits	CareFusion Corp. San Diego, CA	<i>Y-adapter in affected products could develop cracks, potentially resulting in a leak</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	AirLife Infant Breathing Circuits Suspected Lots Recalled: <ul style="list-style-type: none">• Product lot numbers sequentially less than lot # 0000414257• Lot numbers that begin with the letter 'Y' Product Labels Product List	CA , nationwide	Manufactured between June 1, 2010 and February 3, 2012

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/Safety/Recalls/ucm310623.htm>